



Certificate of Analysis

Part No. B2715

Wrought Iron Standard

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone: 01837 54446
Fax: 01837 54544
Web: www.elementalmicroanalysis.com

Certificate Number 120F
Page 1 of 1

% Carbon
Mean = 0.0006
Standard Deviation = ± 0.0002
Expanded Uncertainty = ± 0.0003
(k=2, @ 95% confidence limit)(n=50)

%Sulphur
Mean = 0.0016
Standard Deviation = ± 0.0002
Expanded Uncertainty = ± 0.0004
(k=2, @ 95% confidence)(n=61)

Method of analysis is ASTM E 1019-18 and ARI 033
Primary (NMI)/ISO 17034 Reference Materials Employed:

NIST SRM	2168, 348a
JSS	1205-2, 1201-4
BAM/BCS	270-1, 260/4
NCS	HC11011, HC15002F
ALPHA	AR673-716C, AR673-814H

Notes

The intended use of this reference standard is for the calibration and verification of induction combustion Carbon/Sulphur analysis by infra-red detection as described by ASTM E-1019. The mean analytical values were derived by data sets showing traceability to the above-mentioned primary reference materials and reported in mass fraction. The minimum and typical size for testing was 1g (nominal) per ASTM E1019. The precision values represent the mean value and estimated uncertainty derived from the data sets utilizing ANOVA, ISO Guide 35, and the Guide to Uncertainty Measurement. Metrological traceability is to the SI unit of mass fraction expressed as percent. Refer to your test method for additional uncertainty information. When necessary, professional judgment is applied toward consideration of data and statistical information.

The material used in production of this standard was identified in accordance with ARI 032. The samples for round robin style testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle contains 454g iron chip to be used directly from the bottle with no preparation needed. While unable to determine a definite shelf life this reference should be reviewed 25 years after date of certification. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. This certificate cannot be reproduced except in full.

This Reference Material (RM) is traceable to the above-mentioned reference materials. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use.

Certified July 8, 2020

Elemental Microanalysis Limited