

Certificate of Analysis

(Elemental Composition)

Protein (Casein) Standard OAS
Cat No. B2154 – Certificate No.199622

General

This Organic Analytical Standard (OAS) consists of a homogenous batch of Protein (Casein) standard for use as a routine working micro-analytical standard for the determination of Nitrogen

Certified Values and Uncertainty

The uncertainty in the certified value is expressed as expanded uncertainty, U, at 95% confidence and is calculated in accordance with ISO/IEC17025 according to GUM (Guidelines to Uncertainty in Measurement). Confidence limits include those due to sampling variation, weighing, calibration, and measurement errors. The certified values are based upon the results of 24 to 29 determinations.

The certified value for Nitrogen was determined by elemental analyser calibrated to Cystine 143d from National Institute of Standards and Technology (NIST), Maryland, USA.

<i>Element</i>	<i>Certified value</i> (% w/w)	<i>Uncertainty</i> (+/- %)
Nitrogen	13.32	0.40

Expiration of Certification

The certification of this OAS is valid until 6-Apr-2019 within the measurement uncertainties specified

Storage and use

This OAS should be stored between 20°C to 25°C and should be kept tightly sealed away from light and moisture. It is non-hygroscopic under normal conditions and can be used without preliminary drying.

Certification Information

The technical aspects involved in the preparation, certification and issuance of this (In)Organic Analytical Standard (IAS/OAS) were carried out at Elemental Microanalysis Ltd, Okehampton, Devon EX20 1UB UK, Tel +44 1837 54446, Fax +44 1837 54544, Email info@microanalysis.co.uk.

For and on behalf of
Elemental Microanalysis Ltd

Jon Davies
Technical Manager

