

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis****Part No. B2301****Coal Standard**

Certificate Number 011015  
Page 1 of 1

**Dried Basis Value****Weight Percent Sulphur = 0.57****Expanded Uncertainty = 0.03**

Method

ASTM D 4239-14 and ARI-035

This CRM is traced to N.I.S.T. SRM 2682b, 2692c, 2693, NCS FC28004b

## Notes

The analytical values were derived by a number of data points (n=30) and reported in mass fraction. The precision value represents expanded degree of uncertainty based on errors from analytical assay at a 95% confidence interval (k=2) and may not fit within your testing capabilities. Formal testing procedures should be followed when using this standard; this includes using the *reproducibility* and *repeatability* factors for establishing analytical uncertainty.

The material used in the production of this standard was identified in accordance with ARI 041. The samples used for round robin testing were selected in accordance with ARI 031. The above values relate only to the material used to produce this standard. The analytical samples were dried as per the NMI used or corrected for moisture as per the test method.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. The above values relate only to the material used to produce this standard. This bottle contains 50g, minus 60 mesh (250 micron) coal powder. When kept stored and sealed properly this product has an indefinite shelf life. Once opened, this certificate is valid for two years.

This is a Certified Reference Material (working standard), and is traceable to the above-mentioned NMI standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This CRM is valid for two years from the date of opening

Elemental Microanalysis Limited

Certified June 28, 2016