

Dried basis value

Percent Sulfur (mass fraction) = 2.03

Expanded Uncertainty = ± 0.07

(k=2, @95% confidence limit, n=39)

Method:

ASTM D 4239-18 AND ARI-LAB-616

This CRM is traced to N.I.S.T. SRM: 2683c, 2684c and NCS CRM: FC28011d

Alpha CRM- AR1706-706818, AR177-707812, AR1705-705517, AR1706-706717, AR1705-051020,
AR1707-071222

The intended use of this reference material is for the calibration and quality validation of Sulfur by resistance furnace combustion, infra-red detection analysis as specified by ASTM D4239 or other valid test methods. The analytical values were derived by a number of data sets and reported in mass fraction. The sample size used, and minimum sample size is approximately 300-500mg. The precision value represents the expanded degree of uncertainty based on errors from analytical assay at a 95% confidence level (k=2) utilizing ISO Guide 35, ANOVA, and the Guide to Uncertainty Measurement. Metrological traceability is to the SI derived unit of mass fraction expressed as percent. Formal testing procedures should be followed when using this standard; this includes using the reproducibility and repeatability factors of the method for establishing overall analytical uncertainty. When necessary, professional judgment is applied toward consideration of data and statistical information.

The material used in production of this reference standard was identified in accordance with ARI-LAB-603. The samples for round robin testing were selected in accordance with ARI-LAB-625. The above values relate only to the material used to produce this reference. The analytical samples were dried per the NMI used or corrected for moisture as per the test method. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. This bottle contains 50g, minus 60 mesh (250 micron) coal powder. Values are valid for 15 years from the date of certification. This certificate cannot be reproduced except in full.

This Reference Material (RM) is traceable to the above-mentioned reference materials. For good laboratory practice, it is recommended that all reference materials be verified as fit for purpose prior to use.

Certified on the 7th of June 2023

Elemental Microanalysis Ltd