

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis****Part No. B2308****Coal Standard**

Certificate Number 708716  
Page 1 of 1

**Dried Basis Value**

**Weight Percent Sulphur = 3.03**  
**Expanded Uncertainty = 0.10**

Method  
ASTM D 4239-14 and ARI-035

This CRM is traced to NIST SRM 2684b, 2683c, and NCS FC28012c.

The intended use is for the calibration and validation of sulphur in coal by combustion IR detection. The analytical values were derived by a number of data points (n=40) and reported in mass fraction. The sample size used and minimum sample size is approximately 300-350mg. The precision value represents expanded degree of uncertainty based on errors from analytical assay at a 95% confidence interval (k=2) and may not fit within your testing capabilities. Formal testing procedures should be followed when using this standard; this includes using the *reproducibility* and *repeatability* factors for establishing overall analytical uncertainty.

The material used in the production of this standard was identified in accordance with ARI 041. The samples for round robin testing were selected in accordance with ARI 031. The above values relate only to the material used to produce this standard. The analytical samples were dried as per the NMI used or corrected for moisture as per the test method.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. The above values relate only to the material used to produce this standard. This bottle contains 50g, minus 60 mesh (250 micron) coal powder. When kept stored and sealed properly this product has an indefinite shelf life. Once opened, this certificate is valid for two years.

This is a Certified Reference Material (working standard) and is trace-able to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This CRM is valid for two years from the date of opening

Certified March 1, 2017

Elemental Microanalysis Limited