

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

Certificate of Analysis**Part No. B2330****Lube Oil Standard**

Certificate Number	864708
Date	10 November 2014
Page	1 of 1

Weight Percent Sulphur = 30.00
Standard Deviation = 2.10

Methods

ASTM D-1552
ASTM D-2622
ASTM D-5453-93

This Certified Reference Material was traced to N.I.S.T. SRM 1620b

Notes

The certified mean values were statistically derived from the values reported through Round Robin

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above-mentioned standard. For good laboratory practice it is recommended that all standards be verified prior to use.

This CRM is valid for two years from the date of opening

Elemental Microanalysis Limited