

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis  
Part No. B2333  
High Sulphur Residual Oil  
Standard**

Certificate Number	861694I
Date	17 April 2013
Page	1 of 1

**Weight Percent Sulphur = 5.78  
Standard Deviation = 0.06**

Methods:

ASTM D 4294

This Certified Reference Material was traced to NIST SRM 1620a

**Notes**

Before use, the contents of the bottle should be mixed thoroughly by vigorous shaking. The contents should not be exposed to the air for long periods of time.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is trace-able to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified before use. This CRM is valid for two years from the date of purchase.

Elemental Microanalysis Limited