



Certificate of Analysis
Part No. B2339
Sulphur in Coal Standard

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone: 01837 54446
Fax: 01837 54544
Web: www.elementalmicroanalysis.com

Certificate Number LIG96
Page 1 of 1

Dried Basis Value
Weight Percent Sulphur = 1.18
Standard Deviation = 0.03

Method:
ASTM D 4239 C

NIST standards employed were SRM 2684 & 2685

Notes

Normal ASTM procedures should be employed when using this standard. This includes using the *reproducibility* and *repeatability* factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

EXPIRATION DATE:
THIS RM IS VALID FOR TWO YEARS FROM THE DATE OF OPENING

Elemental Microanalysis Limited