

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

Certificate of Analysis**Part No. B2350****Calcium Oxalate Standard**

Certificate Number	L11Z048
Date	13 April 2015
Page	1 of 1

%Weight Loss @ 200c as H₂O = 12.21

%Weight Loss @ 450c as CO = 18.97

%Weight Loss @ 850c as CO₂ = 29.82

Total % Weight Loss = 61.00

Notes

Samples should be dried for one hour at 105c prior to use

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

Elemental Microanalysis Limited