



Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone: 01837 54446  
Fax: 01837 54544  
Web: [www.elementalmicroanalysis.com](http://www.elementalmicroanalysis.com)

**Certificate of Analysis**  
**Part No. B2351**  
**Calcium Oxalate Standard**

Certificate Number	S06D017
Page	1 of 1

**%Weight Loss @ 200c as H<sub>2</sub>O = 12.21**

**%Weight Loss @ 450c as CO = 18.97**

**%Weight Loss @ 850c as CO<sub>2</sub> = 29.82**

**Total % Weight Loss = 61.00**

**Notes**

Samples should be dried for one hour at 105c prior to use

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

Elemental Microanalysis Limited