

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

**Certificate of Analysis
Part No. B2358
Residual Oil Standard**

Certificate Number 811606
Page 1 of 1

Weight Percent Sulphur = 0.1503
Standard Deviation = 0.0045

Methods
ASTM D-2622
ASTM D-4294
ASTM D-5453-93

This certified reference material was traced to N.I.S.T. SRM 1617a

Notes

The above value represents the mean from the data, reported through Round Robin

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

The material used in the production of this standard was sampled in accordance with ARI 031. The above values relate only to the material used to produce this standard.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (CRM), and is traceable to the above mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

This CRM is valid for two years from the date of purchase

Elemental Microanalysis Limited