

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

**Certificate of Analysis
Part No. B2363
Residual Oil Standard**

Certificate Number	894405
Date	17 April 2013
Page	1 of 1

Residual Oil Standard

**Weight Percent Sulphur = 0.430
Standard Deviation = 0.019**

The above value represents the mean values statistically derived from the data reported from a round robin program.

Methods:
ASTM D 2622.
ASTM D 5453.

This Certified Reference Material was traced to N.I.S.T SRM 1624c.

Notes

Before use, the contents of the bottle should be mixed thoroughly by vigorous shaking of the bottle. The contents should not be exposed to the air for long periods of time.

Normal ASTM procedures should be employed when using this standard. This includes using the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

This is a Certified Reference Material (CRM) and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

For and on behalf of
Elemental Microanalysis Limited

This CRM is valid for two years from the date of opening.