

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis  
Part No. B2370  
Residual Oil Standard**

Certificate Number	818411
Date	17 April 2013
Page	1 of 1

**Weight Percent Sulphur = 1.49  
Standard Deviation = 0.06**

High Purity Materials/Standard used:

N.I.S.T. SRM 1622b

Method used for analysis:

ASTM D-4294

This standard was produced gravimetrically using high purity materials, with balances calibrated and checked by precision NIST traceable weights, and verified using the above ASTM method. The precision value represents the 95% confidence limit ( $k=2$ ) derived from analysis and errors from weighing. This includes the reproducibility and repeatability factors for the ASTM method you wish to employ.

Notes:

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

This is a Certified Reference Material (CRM). For good laboratory practice it is recommended that all standards be verified prior to use.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This CRM is valid for two years from the date of opening

Elemental Microanalysis Limited

Certified April 20, 2011