

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis  
Part No. B2374  
Residual Oil Standard**

Certificate Number 898901  
Date 17 April 2013  
Page 1 of 1

**Weight Percent Sulphur = 3.46  
Standard Deviation = 0.04**

The certified mean values were statistically derived from the values reported through Round Robin

**Methods**

ASTM D-4294

**Notes**

Traced to N.I.S.T. SRM 1620b & 1622e

Before use, the contents of the bottle should be mixed through vigorous shaking. The contents should not be exposed to air for a lengthy period of time.

Normal ASTM procedures should be employed when using this standard. This includes the reproducibility and repeatability factors. Use the reproducibility factor when establishing the first set of data, and then establish an internal repeatability factor as your guideline.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This CRM is valid for two years from the date of purchase

Elemental Microanalysis Limited