

Analytical Results

% Oxygen

Value = 0.124

Expanded Uncertainty = ± 0.015

Method & Detection: Inert Gas Fusion

$k \approx 2$ (95% confidence)

% Nitrogen

Value = 0.245

Expanded Uncertainty = ± 0.020

Method & Detection: Inert Gas Fusion

$k \approx 2$ (95% confidence)

Primary Reference Standards Used:

Methods Employed:

JCGM 100:2008; Evaluation of Measurement Data – Guide to the Expression of Uncertainty in Measurement; (GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (JCGM) (2008); available at https://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed February 2025)

JCGM 101:2008; Evaluation of Measurement Data – Supplement 1 to the Guide for the Expression of Uncertainty in Measurement; Propagation Distributions Using a Monte Carlo Method; Joint Committee for guides in Metrology (JCGM) (2008); available at

**The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the verification of calibration of inert gas fusion for the determination of oxygen and nitrogen.

The minimum sample size to perform this intended use is no less than 1 pin.

This certification is valid for 20 years from the initial certification date as seen below, provided the RM is handled and stored in accordance with the instructions given in this certificate.

This bottle contains 50g of Steel Pins of nominal mass 0.5 g. to be used per the test method you follow. Keep sealed tightly and store under normal laboratory conditions. This product requires no preparation prior to use.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 5th of March 2026

Elemental Microanalysis Ltd