

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis  
Part No. B2512  
Carbon & Sulphur Pin Standard**

Certificate Number 814G  
Page 1 of 1

**% Carbon****Mean = 0.016****One Sigma Standard Deviation = +/- 0.002****Expanded Uncertainty = +/- 0.004****(k=2, 95% confidence)****%Sulphur****Mean = 0.262****One Sigma Standard Deviation = +/- 0.007****Expanded Uncertainty = +/- 0.014****(k=2, 95% confidence)**

Method of analysis is ASTM E 1019-11 and ARI 033

Primary (NMI) Standards Employed:

NIST SRM 2159, 129c, 133b, 101g

BAM 085-1, 284-2

**Notes**

The mean analytical values shown are derived by 5 data sets (n=50) showing traceability to the above mentioned reference standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed.

The material used in production of this standard was sampled in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle consists of 250g, 0.5g pins (nominal weight) to be used directly from the bottle with no preparation necessary. This product has an indefinite shelf life. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (Working Standard), and is traceable to the above-mentioned NMI standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Elemental Microanalysis Limited

Certified December 9, 2014