

**Certificate of Analysis
Part No. B2518
Carbon & Sulphur Pin Standard**

Certificate Number 119E
Page 1 of 1

% Carbon
Mean = 0.134
Standard Deviation = +/- 0.002
Expanded Uncertainty = +/- 0.006
(k=2, @95% confidence)(n=39)

%Sulphur
Mean = 0.229
Standard Deviation = +/- 0.007
Expanded Uncertainty = +/- 0.015
(k=2, @95% confidence)(n=36)

Method of analysis is ASTM E1019-18 and ARI 033

Primary (NMI)/Guide 34 Reference Standards employed:

NIST SRM	133b, 129c, 19h
JSS	154-14
EURO	CRM232/2, 085-1, 058-2, 286-1, 079-1
ALPHA	AR888-217A, AR891-817E, AR896-1114A, AR947-1114B

Notes

The intended use of this reference standard is for the calibration and verification of induction combustion Carbon/Sulphur analysis by infra-red detection as described by ASTM E1019. The mean analytical values were derived by separate data sets showing traceability to the above-mentioned primary reference standards and reported in mass fraction. The minimum and typical size for testing was 1g (1 pin) per ASTM E1019. The precision values represent the mean value and estimated uncertainty derived from the data sets. Refer to your test method for additional uncertainty information. When necessary, professional judgment is applied toward consideration of data and statistical information.

The material used in production of this standard was identified in accordance with ARI 032. The samples for round robin style testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle contains 454g, 1g pins (nominal weight), to be used directly from the bottle with no preparation needed. While unable to determine a definite shelf life this reference should be reviewed 25 years after date of certification. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event, shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. This is a Reference Material is traceable to the above-mentioned reference materials. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use.

Certified 24th April 2019

Elemental Microanalysis Limited