

Elemental Microanalysis Ltd 1 Hameldown Road Okehampton EX20 1UB United Kingdom Telephone: 01837 54446 Fax: 01837 54544 Web: www.elementalmicroanalysis.com

% Carbon

Mean = 0.131Standard Deviation = ± 0.001 Expanded Uncertainty = ± 0.004 (k=2, @95% confidence limit) (n=30)

Certificate of Analysis Part No. B2518 Steel Pin Standard

Certificate Number: 722D Page 1 of 1

% Sulphur Mean = 0.228 Standard Deviation = ± 0.005 Expanded Uncertainty = ± 0.012 (k=2, @95% confidence limit) (n=30)

 Method of Analysis is ASTM E1019-18 and ARI-LAB-621

 Primary (NMI)/ISO17034 Reference Standards Employed:

 NIST SRM
 129c, 343a, 133b, 293

 BAM
 286-1, 079-1

 NCS
 HC11111

 ALPHA – AR888-217A, AR896-1114A, AR888-8E8539, AR883-216B, AR890-217D, AR951-918C,

 AR956-113A

The intended use of this reference standard is for the calibration and verification of induction combustion Carbon/Sulphur analysis by infra-red detection as described by ASTM E1019. The mean analytical values were derived by data sets showing traceability to the above-mentioned primary reference standards and reported in mass fraction. The minimum and typical size for testing was 1g (1 pin). The precision values represent the mean value and estimated uncertainty derived from the data sets utilizing ANOVA, ISO Guide 35, and the Guide to Uncertainty Measurement. Metrological traceability is to the SI unit of mass fraction expressed as percent. Refer to your test method for additional uncertainty information. When necessary, professional judgment is applied toward consideration of data and statistical information.

The material used in production of this standard was identified in accordance with ARI-LAB-603. The samples for analysis were selected in accordance with ARI-LAB-625. The above values relate only to the material used to produce this standard. This bottle contains 454g, 1g pins (nominal weight), to be used directly from the bottle with no preparation needed. While unable to determine a definite shelf life this reference should be reviewed 25 years after date of certification. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event, shall Elemental Microanalysis Ltd be liable for incidental or consequential damages. This certificate cannot be reproduced except in full.

This Reference Material is traceable to the above-mentioned reference materials. For good laboratory practice, it is recommended that all standards be verified as fit for purpose prior to use.

Certified on the 21st October 2022

Elemental Microanalysis Ltd