

Analytical Results	
% Carbon Value = 0.276 Expanded Uncertainty = 0.005 Method & Detection = Combustion/IR n = 40 k \approx 2 (95% confidence)	% Sulfur Value = 0.0227 Expanded Uncertainty = 0.0019 Method & Detection = Combustion/IR n = 40 k \approx (95% confidence)
Primary Reference Standards Used: NIST SRM 293, 100b, 123c, 19h, 106b NCS HC 11326, HC 11325	
Methods Employed: ASTM E1019 – Standard Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel, Iron, Nickel, and Cobalt Alloys by Various Combustion and Inert Gas Fusion Techniques	

**The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the verification and calibration of induction furnace combustion and other appropriate analysis methods for the determination of carbon and sulfur.

The minimum sample size to perform this intended use is dependent upon the test method and instrumentation used. It is recommended that no less than 1 pin of CRM material be used for destructive test methods.

The Period of Validity for this RM is 15 years after opening, if handling and storage instructions are followed, and should be reviewed 15 years after the date below.

This bottle contains 454g of nominal 1.0 g pins to be used per the test method you follow. Keep sealed tightly and store under normal laboratory conditions.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 9th of December 2024

Elemental Microanalysis Ltd