

Elemental Microanalysis Ltd 1 Hameldown Road Okehampton EX20 1UB United Kingdom Telephone: 01837 54446 Fax: 01837 54544 Web: www.elementalmicroanalysis.com

Certificate of Analysis Part No. B2362 Hydrogen in Steel Pin Standard

RM Doc Number: 250415

Page **1** of **1**

Analytical Results

% Hydrogen

Value = 0.00061 Expanded Uncertainty = 0.00014 Method & Detection = Inert gas fusion $k \approx 2 \; (95\% \; confidence)$

Methods Employed:

JCGM 100:2008; Evaluation of Measurement Data – Guide to the Expression of Uncertainty in Measurement; (GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (JCGM) (2008); available at https://www.bipm.org/utils/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed February 2025) JCGM 101:2008; Evaluation of Measurement Data – Supplement 1 to the Guide for the Expression of Uncertainty in Measurement; Propagation Distributions Using a Monte Carlo Method; Joint Committee for guides in Metrology (JCGM) (2008); available at

https://www.bipm.org/utils/common/documents/jcgm/JCGM_100_2008.pdf (accessed February 2025).

The intended use of this Reference Material (RM) is for the verification and calibration of inert gas fusion for the determination of hydrogen.

The minimum sample size to perform this intended use is no less than 1 pin of RM material be used for destructive test methods.

The Period of Validity for this RM is 20 years from the initial certification date, provided the RM is handled and stored in accordance with the instructions given in this certificate.

This bottle contains 100g of pins of nominal mass 1.0 g to be used per the test method you follow. Keep sealed tightly and store under normal laboratory conditions.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 21st of May 2025

Elemental Microanalysis Ltd

^{*}The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.