Elemental Microanalysis Limited

Okehampton Business Park Exeter Road Okehampton Devon EX20 1UB Telephone 01837 54446/7 Fax 01837 54544

Certificate of Analysis Part No. B2700 Cast Iron Standard

Certificate Number 614B Page 1 of 1

% Carbon
Mean = 0.97
One Sigma Standard Deviation = +/- 0.01
0.0015

%Sulphur
Mean = 0.0102
One sigma Standard Deviation = +/-

Expanded Uncertainty =+/- 0.02 (k=2, 95%confidence)

Expanded Uncertainty =+/- 0.0030 (k=2, 95%confidence)

Method of analysis is ASTM E 1019-11 and ARI 033.

Primary (NMI) Standards Employed: NIST 337a, 345a, 155 JSS 602-10 EURO 035-1, 226-1, 92

Notes

The mean analytical values were derived by 6 data sets, (n=60), showing trace-ability to the above mentioned NMI standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed.

The material used in production of this standard was sampled in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle contains 150g, cast iron powder, to be used directly from the bottle with no preparation needed. This product has an indefinite shelf life.

Keep sealed and store in a cool dry place

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a certified reference material (working standard) and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all secondary standards be verified prior to use

Elemental Microanalysis Limited

Certified June 30, 2014