

Analytical Results

% Carbon

Mean = 3.08

Standard Deviation = ± 0.03

Expanded Uncertainty = ± 0.07

(k=2, @95% confidence. n=31)

% Sulfur

Mean = 0.108

Standard Deviation = 0.004

Expanded Uncertainty = ± 0.010

(k=2, @95% confidence. n=34)

Primary (NMI)/Guide 34/ISO 17034 Standards Employed:

NIST	4L, 6g, 107c
NCS	HC11007
JSS	120-1
BAS	484-1
CTIF	FB12, FB 10-1

Method of Analysis is ASTM E 1019-18

**The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the calibration and continued quality validation of Carbon and Sulfur in induction combustion, IR detection systems as described in ASTM E1019.

The minimum sample size to perform this intended use is subject to the ASTM test method and detection capability of your analyzer. The analytical sample test size used was 0.3-0.5g nominal as suggested by the instrument manufacturer(s).

The Period of Validity for this RM is not able to be determined and should be reviewed 25 years after the date below.

This bottle contains 150g of cast iron shot to be used per the test method you follow. Keep sealed tightly and store under normal laboratory conditions.

Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 19th of May 2020.

Elemental Microanalysis Ltd