

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

Certificate of Analysis**Part No. B2705****Iron Standard**

Certificate Number 1857
Date 18 April 2013
Page 1 of 1

Oxygen and Nitrogen

& Carbon
3.83 +/- 0.03

% Sulphur
0.016 +/- 0.001

Round Robin Summary.

Lab	Carbon %	Sulphur %	Method
1	3.86	0.016	1
2	3.83	0.016	1
3	3.80	0.016	1
4	3.84	0.014	1
5	3.82	0.015	1

Method of analysis is
Combustion /IR Detection ASTM E 1019

The following standards were employed:

NIST SRM 122h, 4k, 338, 892

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

For and on behalf of
Elemental Microanalysis Limited

Stuart C. Carter
Technical Director

Lot number 03495