

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

Certificate of Analysis**Part No. B2710****Cast Iron Standard**

Certificate Number 1012A
Page 1 of 1

% Carbon**Mean =3.26****One Sigma Standard Deviation = +/- 0.04****Expanded Uncertainty = +/- 0.08****(k=2, 95% confidence)****% Sulphur****Mean = 0.021****One Sigma Standard Deviation = +/- 0.001****Expanded Uncertainty = +/-0.002****(k=2, 95% confidence)**

Method of analysis is ASTM E1019-11 and ARI 033

Primary (NMI) Standards Employed:

NIST	338
CKD	227
IPT	66
JSS	111-12

Notes

The mean analytical values shown are derived by 5 data sets (n=50), showing trace-ability to the above mentioned NMI standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed.

The material used in production of this standard was sampled in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle consists of 150g, cast iron powder, to be used directly from the bottle with no preparation needed. This standard has an indefinite shelf life. Keep sealed and store in a cool dry place

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (Working Standard), and is traceable to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Elemental Microanalysis Limited

Certified November 14, 2012