

**Elemental Microanalysis Limited**

Okehampton Business Park  
Exeter Road  
Okehampton  
Devon EX20 1UB  
Telephone 01837 54446/7  
Fax 01837 54544

**Certificate of Analysis  
Part No. B2715  
Wrought Iron Standard**

Certificate Number 716C  
Page 1 of 1

**% Carbon****Mean = 0.0006****One Sigma Standard Deviation = +/- 0.0002****Expanded Uncertainty = +/- 0.0004****(k=2, 95% confidence)(n=50)****%Sulphur****Mean = 0.0011****One Sigma Standard Deviation = +/- 0.0002****Expanded Uncertainty = +/- 0.0004****(k=2, 95% confidence)(n=49)**

Method of analysis is ASTM E 1019-11 and ARI 033

Primary (NMI) Standards Employed:

NIST	348a, 2168, 131g
NCS	HC11011, HC15002F
JSS	003-6, 1201-4, 244-11, 1205-2

**Notes**

The mean analytical values shown are derived by data sets showing trace-ability to the above mentioned reference standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for additional uncertainty information.

The material used in production of this standard was sampled in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle consists of 454g, granular chips, to be used directly from the bottle with no preparation needed. This product has an indefinite shelf life. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Material (working standard) and is trace-able to the above-mentioned standards. For good laboratory practice it is recommended that all standards be verified prior to use.

Elemental Microanalysis Limited

Certified September 13, 2016