

Elemental Microanalysis Limited

Okehampton Business Park
Exeter Road
Okehampton
Devon EX20 1UB
Telephone 01837 54446/7
Fax 01837 54544

**Certificate of Analysis
Part No. B2764
Stainless Steel Chip Standard**

Certificate Number 814B
Date 03 March 2015
Page 1 of 1

% Carbon**Mean = 0.121****One Sigma Standard Deviation = 0.003****Expanded Uncertainty = 0.006****(k=2, 95% confidence) n=40****% Sulphur****Mean = 0.0221****One Sigma Standard Deviation = 0.0009****Expanded Uncertainty = 0.0018****(k=2, 95% confidence) n=40****% Nitrogen****Mean = 0.036****One Sigma Standard Deviation = 0.001****Expanded Uncertainty = 0.002****(k=2, 95% confidence) n=40****% Oxygen = (0.0083) reference only**

Method of analysis is
ASTM E 1019-11, ARI 033 and ARI 034

Primary (NMI) Standards employed:
NIST SRM 343a, 293, 73c, 13g, 100b, 16f, 133b, 129c
IPT 26
JSS 610-10
EURO 227-1, 226-1

Notes

The mean analytical values were derived by data sets showing trace-ability to the above mentioned primary standards, and reported in mass fraction. The precision values represent the estimated uncertainty derived from the data sets and may not represent your testing capabilities. Refer to your test method for the expanded method derived uncertainty if needed

The material used in production of this standard was evaluated and determined in accordance with ARI 032. The samples used for round robin testing were selected in accordance with ARI 014. The above values relate only to the material used to produce this standard. This bottle consists of 150g, clean chips, to be used directly from the bottle with no preparation needed. This product has an indefinite shelf life. Keep sealed and store under normal laboratory conditions.

Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd be liable for incidental or consequential damages.

This is a Certified Reference Standard (working standard), and is traceable the above-mentioned references. For good laboratory practice it is recommended that all standards be verified prior to use

Elemental Microanalysis Limited

Certified December 17, 2014