

Analytical Results

% Sulfur

Value = 2.10

Expanded Uncertainty = ± 0.14

Method & Detection: X-ray Fluorescence Spectroscopy

$k \approx 2$ (95% confidence)

Methods Employed:

JCGM 100:2008; Evaluation of Measurement Data – Guide to the Expression of Uncertainty in Measurement;(GUM 1995 with Minor Corrections), Joint Committee for Guides in Metrology (JCGM) (2008); available at https://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed February 2025)

JCGM 101:2008; Evaluation of Measurement Data – Supplement 1 to the Guide for the Expression of Uncertainty in Measurement; Propagation Distributions Using a Monte Carlo Method; Joint Committee for guides in Metrology (JCGM) (2008); available at https://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008.pdf (accessed February 2025).

**The analytical results above are provided by an accredited reference material manufacturer with a current certification in ISO 17025 and 17034.*

The intended use of this Reference Material (RM) is for the verification of x-ray fluorescence spectroscopy for the determination of Sulfur. The minimum sample size to perform this intended use is 2 mL of material be used for non-destructive test methods.

The Period of Validity for this RM is 5 years from the initial certified date, provided the RM is handled and stored in accordance with the instructions given in this certificate. This product requires no preparation prior to use. Bottles of liquid should be kept sealed tight and stored in a cool, dry location.

This bottle contains 100 mL of liquid to be used per the test method you follow. Refer to your test methods and or manufacturer manual for expanded uncertainties, repeatability/reproducibility factors.

For good laboratory practice, we recommend that all reference materials be verified as fit for purpose prior to use. Remedies for any claimed defect in this product will be limited to product replacement or refund of the purchase price. In no event shall Elemental Microanalysis Ltd. be liable for incidental or consequential damages.

Certified on the 5th of August 2025

Elemental Microanalysis Ltd