

Technical note

B1297 Sicapent[®] with Indicator 30g

Due to the extremely high hygroscopic nature of this product and the sensitivity of the indicator, traces of atmospheric moisture may be absorbed during the bottling process causing a pale green colouration.

This trace absorption does not signify that the bottle has leaked or that the product will be ineffective in operation.



 Elemental
Microanalysis

These products offer the same high quality and value for money that is typical of Elemental Microanalysis' entire product range.

Available to order from our [ISO 9001](#) factory online, by email, telephone or fax. Elemental Microanalysis offers extensive knowledge of analysis, over 40 years of experience and a worldwide reputation for excellence.

All Elemental Microanalysis products are guaranteed to be of high quality and are deemed to be suitable alternatives in the stated instruments. All trademarks acknowledged.

Elemental Microanalysis Ltd

1 Hameldown Road, Okehampton, EX20 1UB, UK
t. +44(0)1837 54446 • www.elementalmicroanalysis.com •
enquiries@microanalysis.co.uk



Available from your local Dealer:

